Version 210712

Analysenzertifikat
Certificate of Analysis

NGS DNA Purification MagBeads	Kat. Nr.: / Cat. No.:	S5352
NGS DNA Purification MagBeads	Lot / Contr. No.:	NK2357B13W01

Parameter **Parameter**	**Methode** **Method**	**Spezifikation** **Specification**	**Ergebnis** **Result**
Library yield:	The control yield of blank library was no more than 30ng. The library yield is not less than 1000ng. The repeat error of Complex hole library yield was not more than ±20%, and the average yield difference between the test group and the control group was not more than ±20%.	In range	Pass
Library size:	The library has no linker dimer and no large segment peak, and the peak size range between the test group and the control group is 200-400 bp.	In range	Pass
Nächste QC: Next retest:		08/2024	

Versand: ungekühlt. / Lagerung: bei +2°C bis +8°C. Nicht einfrieren!
Shipment: not cooled. / Storage: at +2°C to +8°C. Don't freeze!

Verwendung
Laborreagenz. Nur für Forschungszwecke zugelassen.
Nicht für die Verwendung an/in Mensch oder Tier. Keine Anwendung als Diagnostika. Nicht für den Verzehr.

Suitability
For research usage only.
Not approved for human or animal diagnostic or therapeutic procedures. Not for consumption.

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

Genaxxon bioscience GmbH

Datum/date: 24.02.2024

Dr. Norbert Tröndle

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This document has been generated by computer and does not contain a signature.

Genaxxon bioscience GmbH - Söflinger Str. 70 - D-89077 Ulm - Tel.: +731 3608 123 - Fax: +731 3608 962